

Commission number: [REDACTED]



## CONTENTS

Number of ordered detectors: 1

Product: Radtrak<sup>2</sup> Long-Term detectors

Measurement period: 90-365 days

Radtrak<sup>2</sup> detectors should be returned for analysis within 3 years of Radonova delivery date.

## INSTRUCTION FOR LONG-TERM RADON MEASUREMENT

[www.radonline.com](http://www.radonline.com) Commission nbr: [REDACTED] Password: [REDACTED]

- 1 Read these instructions in their entirety. The detector(s) should be placed accordingly, with the measurement period being at least 90 days, not exceeding one year. The measurement begins when the radon-proof plastic bag is opened. Immediately place device.
- 2 Go to [www.radonline.com](http://www.radonline.com) and log in to 'Commission Management' using your commission number and password above. Fill in measurement information, including specific detector location, detector number, and start date. The detector(s) should remain deployed in the same location until they are ready to be returned.
- 3 Upon completion of test, log in to the website and provide the measurement end date. Confirm your location information and click Save. Package up your detector(s) and return immediately to the laboratory address below. Retain this document for your records.
- 4 Upon receipt your detectors will be analyzed. Measurement Reports are typically available within 5-10 business days. The report can be emailed to the email address provided, and may be retrievable from the website as well.

## RECOMMENDED DETECTOR PLACEMENT

- ✓ The EPA recommends the test kit be placed in the lowest, lived-in level of the home (i.e. a frequently used basement).
- ✓ It should be put in a room that is regularly used such as a living room, den, or bedroom, but not in a kitchen or bathroom.
- ✓ At least 20" above the floor
- ✓ At least 3' from windows or exterior doors
- ✓ At least 4" from objects or outside walls
- ✗ Do not place on any mantels or other natural stone surfaces

**Please note!** Some states or organizations may have differing recommendations. If you have a question regarding placement, contact your state radon agency. The Radtrak<sup>2</sup>® radon detector is manufactured and analyzed by Radonova Laboratories AB, the global leader in radon measurements. This device is designed for long-term testing per EPA protocol, or 90 days to one year. Long-term detectors provide a more reliable representation of your average radon levels, and are less vulnerable to under/over-reporting due to fluctuations in radon levels.

**Return detector(s) to:**  
Radonova, Inc.  
900 Oakmont Lane Suite 207, Westmont, IL 60559

**If you have any questions regarding the measurement performance contact**  
Email: [help@radonova.com](mailto:help@radonova.com)  
Telephone: 331.814.2200

Use this Data Sheet and send in with detectors ONLY if you are not able to enter data into [www.radonline.com](http://www.radonline.com), as instructed on page 1. Using this form in place of web data entry will delay your reporting.

Name:.....  
 Address:.....  
 .....  
 City:.....  
 State:..... County:.....  
 Zip code:.....  
 Building number:.....

Radtrak<sup>2</sup> Long-Term Radon Test Dwelling  
 4730540

+  
**Important!**  
 Email: .....

**DETECTOR AND COMMISSION INFORMATION**

Fill in detector numbers (bottom of the detector)	Specific detector location	Bed room	Living room	Other living area	Non-living area	Floor detector was placed	Use this character B= Basement 0= Ground Floor 1= First Floor 2= Second Floor 3= Third Floor 4= Fourth Floor 5= Fifth Floor 6= Sixth Floor 7= Seventh Floor 8= Eighth Floor 9= Ninth Floor X= 10+ Floor
1		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
2		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
3		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
4		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		

Start Date: [M M D D Y Y]      End Date: [M M D D Y Y]

**BUILDING INFORMATION**

<b>Building Type</b> <input type="checkbox"/> SF Detached <input type="checkbox"/> Condominium <input type="checkbox"/> Apartment <input type="checkbox"/> Townhouse <input type="checkbox"/> Duplex <input type="checkbox"/> Other	<b>Year Built</b> [ ] [ ] [ ] [ ]	<b>Foundation Type</b> <input type="checkbox"/> Basement <input type="checkbox"/> Bi-Level <input type="checkbox"/> Slab On Grade <input type="checkbox"/> Crawl Space <input type="checkbox"/> Pillar/Above grade	<b>Purpose of test</b> <input type="checkbox"/> Primary Screening <input type="checkbox"/> Real Estate Transaction <input type="checkbox"/> Follow Up <input type="checkbox"/> Post Mitigation
---	--------------------------------------	---	--

Printed name: [.....]

Test data given by (signature): .....

If the measurement was performed by a certified measurement specialist, certification license no:.....

Commission number: [REDACTED]

Password: [REDACTED]

RC004LN - V1.50 / 2017-05-04 / J.O. / L.B.

DISCLAIMER - Radonova, Inc. makes no warranty of any kind, express or implied, as regard to the use, operation or analysis of any Radonova, Inc. monitor. Radonova, Inc. specifically disclaims implied warranties of merchantability and fitness for a particular purpose. Radonova, Inc. is not responsible for any damage, including consequential damages, to persons or property resulting from the use of the monitor or the resulting data.