



6946191



Tracy [REDACTED]
7165 [REDACTED]
EGG HARBOR WI 54209

CONTENTSNo. of ordered detectors: **3**Product: **Radtrak³®**

Order ID:

Account: **TRACY [REDACTED]**Measurement period: **90 - 365 days**Deploy By: **Date on Detector****LOGIN CREDENTIALS**Commission number: **6946191**Password: **C8Lb4L2****Radtrak³® | Long-Term Radon Test for Dwellings****INSTRUCTION FOR LONG-TERM RADON MEASUREMENT**

1 Read these instructions in their entirety. The detector(s) should be placed accordingly, with the measurement period being at least 90 days, not exceeding 365 days. The measurement begins when the radon-proof plastic bag is opened. Immediately place device. If the heating season occurs for at least half of the year, then at least 50% the long-term test should be conducted during the heating season.

2 Go to www.radonline.com and use your commission number and password above to log into **MyPages**. Fill in measurement information, including specific detector location, detector number, and start date. The detector(s) should remain deployed in the same location until they are ready to be returned.

3 Upon completion of test, log back into **MyPages** and provide the measurement end date and time. Start date and end date are required to calculate a result. Confirm your location information and Save. Registration must be completed to receive report.

4 Package up and return detector(s) immediately to the laboratory address below. Write down your detector serial number(s) and retain this document for your records. When returning detector(s), a shipping service (USPS, FedEx, etc) with tracking is recommended. Upon receipt your detectors will be analyzed. Measurement reports are typically available within 5-10 business days. The report can be emailed to the address provided and may be retrievable from the website as well.

RECOMMENDED DETECTOR PLACEMENT

- At least 20" above the floor
- At least 3' from windows or exterior doors
- At least 4" from objects or outside walls
- **Do not place on any mantels or other natural stone**

Please note! Some states or organizations may have differing recommendations. If you have a question regarding placement, contact your state radon agency.

- The EPA recommends the test kit be placed in the lowest, lived-in level of the home (i.e. a frequently used basement)
- It should be put in a room that is regularly used such as a living room, den, or bedroom, **but not in a kitchen or bathroom**

The Radtrak³®radon detector is manufactured and analyzed by Radonova Laboratories, the global leader in radon measurements. This device is designed for long-term testing per EPA protocol, or 90 days to one year. Long-term detectors provide a more reliable representation of your average radon levels, and are less vulnerable to under/over-reporting due to fluctuations in radon levels.

RETURN DETECTOR(S) TO

Radonova Inc.

1 East 22nd Street, Suite 200, Lombard, IL 60148

DISCLAIMER –Radonova Inc.. makes no warranty of any kind, express or implied, as regard to the use, operation or analysis of any Radonova Inc.. monitor. Radonova Inc.. specifically disclaims implied warranties of merchantability and fitness for a particular purpose. Radonova Inc.. is not responsible for any damage, including consequential damages, to persons or property resulting from the use of the monitor or the resulting data.

QUESTIONS REGARDING THE MEASUREMENTEmail: help@radonova.com

Telephone: 331.814.2200