



CONTENTS

No. of ordered detectors: 1

Product: Radtrak³[®]

Order ID:

Account: KANSAS STATE UNIVERSITY

Measurement period: 90 - 365 days

Deploy By: Date on Detector

LOGIN CREDENTIALS

Commission number: XXXXXXXX

Password: XXXXXXXX

Radtrak³[®] | Long-Term Radon Test for Dwellings

INSTRUCTION FOR LONG-TERM RADON MEASUREMENT

- 1Read these instructions in their entirety. The detector(s) should be placed accordingly, with the measurement period being at least 90 days, not exceeding 365 days. The measurement begins when the radon-proof plastic bag is opened. Immediately place device. If the heating season occurs for at least half of the year, then at least 50% the long-term test should be conducted during the heating season.
- 2Go to www.radonline.com and use your commission number and password above to log into **MyPages**. Fill in measurement information, including specific detector location, detector number, and start date. The detector(s) should remain deployed in the same location until they are ready to be returned.
- 3Upon completion of test, log back into **MyPages** and provide the measurement end date and time. Start date and end date are required to calculate a result. Confirm your location information and Save. Registration must be completed to receive report.
- 4Package up and return detector(s) immediately to the laboratory address below. Write down your detector serial number(s) and retain this document for your records. When returning detector(s), a shipping service (USPS, FedEx, etc) with tracking is recommended. Upon receipt your detectors will be analyzed. Measurement reports are typically available within 5-10 business days. The report can be emailed to the address provided and may be retrievable from the website as well.

RECOMMENDED DETECTOR PLACEMENT

- At least 20" above the floor
 - At least 3' from windows or exterior doors
 - At least 4" from objects or outside walls
 - Do not place on any mantels or other natural stone
- Please note!** Some states or organizations may have differing recommendations. If you have a question regarding placement, contact your state radon agency.

 - The EPA recommends the test kit be placed in the lowest, lived-in level of the home (i.e. a frequently used basement)
 - It should be put in a room that is regularly used such as a living room, den, or bedroom, but not in a kitchen or bathroom

The Radtrak³[®]radon detector is manufactured and analyzed by Radonova Laboratories, the global leader in radon measurements. This device is designed for long-term testing per EPA protocol, or 90 days to one year. Long-term detectors provide a more reliable representation of your average radon levels, and are less vulnerable to under/over-reporting due to fluctuations in radon levels.